

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product
Produit

Power supply for building-in

Name and address of the applicant
Nom et adresse du demandeurEGSTON System Electronics Eggenburg GmbH
Grafenberger Strasse 37, A-3730 Eggenburg, AustriaName and address of the manufacturer
Nom et adresse du fabricantEGSTON System Electronics Eggenburg GmbH
Grafenberger Strasse 37, A-3730 Eggenburg, AustriaName and address of the factory
Nom et adresse de l'usineEGSTON System Electronic spol.s r.o.
Průmyslová 20, CZ-670 15 Znojmo, Czech Republic*Note: When more than one factory, please report on page 2*
Note: Lorsque il y a plus d'une usine, veuillez utiliser la 2ème page Additional Information on page 2Ratings and principal characteristics
Valeurs nominales et caractéristiques principalesInput: 100-240 Vac; 50-60 Hz; 340 mA
See page 2.Trademark (if any)
Marque de fabrique (si elle existe)

EGSTON

Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeur

/

Model / Type Ref.
Ref. De typeN1hxyW3 18
See page 2.Additional information (if necessary may also be reported
on page 2)
Les informations complémentaires (si nécessaire, peuvent
être indiqués sur la 2^{ème} pageThe risk management requirements of the standard were not
addressed. Additional Information on page 2A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à laIEC 60601-1:2005 (3rd Ed.) + A1:2012As shown in the Test Report Ref. No. which forms
part of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

T223-0596/16, dated 2016-09-29

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**SIQ Ljubljana
Tržaška cesta 2, SI-1000 Ljubljana, Slovenia
T +386 1 4778 100, F +386 1 4778 444, info@siq.si, www.siq.si
Product Certification Body is accredited by Slovenian Accreditation, Reg. No.: CP-001

Date: 2016-09-29

Signature: Bojan Pečavar

Name and address of the factory:

- 1) EGSTON System Electronic spol.s r.o.
Průmyslová 20, CZ-670 15 Znojmo, Czech Republic
- 2) EGSTON Electronics Zhuhai Ltd.
3/F, No. 3 Pinggong No. 1 Road, Nanping Technology Ind. Park, Zhuhai City, China

Ratings and principal characteristics:

Input:

100-240 Vac; 50-60 Hz; 340 mA

Output voltage: 5,0 Vdc to 24,0 Vdc

Output current: Max. 2,5 A

Output power: Max. 18 W

Model / Type Ref.:

N1hxyW3 18

“x” can be:

“F” for cable version

“P” for PCB version

“K” for terminal version

“I” for insulated cable version

“y” can be S (for IT, Household and Medical) or M (for Medical)

Additional information (if necessary)

Information complémentaire (si nécessaire)

Date: 2016-09-29

Signature: Bojan Pečavar

