

# Test report

**Number:** T251-0824/20 **Project file:** C20202804  
**Date:** 2021-01-11  
**Pages:** 89

**Product:** Switch Mode Power Supply

**Type reference:** E2xFyW3\_z 60; where  
'x' can be x (no special option) or O (Open Frame Version) or D (Desktop), and  
'y' can be x or y (no special options) or S (Standard) or H (Household) or M (Medical), and  
'\_z' can be blank or \_z (no special option) or \_E (Protective earth connected to output)  
See page 6.

**Ratings:** See page 6.

**Trademark:** **Pulse** or **EGSTON**

**Applicant:** EGSTON System Electronics Eggenburg GmbH  
Grafenberger Strasse 37, AT-3730 Eggenburg, Austria

**Manufacturer:** EGSTON System Electronics Eggenburg GmbH  
Grafenberger Strasse 37, AT-3730 Eggenburg, Austria

**Place of manufacture:** EGSTON System Electronic spol.s.r.o.  
Průmyslova 20, CZ-670 15 Znojmo, Czech Republic

## Summary of testing


**Testing method:** EN 60601-1-2:2015, EN 55032:2015 + A11:2020, EN 55024:2010,  
EN 61000-3-2:2014, EN 61000-3-3:2013,  
EN 61000-6-3:2007 + A1:2011, EN 61000-6-1:2007

**Testing location:** SIQ Ljubljana, Trpinčeva ulica 37 A, SI-1000 Ljubljana, Slovenia (Initial testing)  
SIQ Ljubljana, Mašera-Spasičeva ulica 10, SI-1000 Ljubljana, Slovenia (other Revisions)

**Remarks:** Date of receipt of test items: see page 4.  
Number of items tested: see page 4.  
Date of performance of tests: see page 4.  
The test results presented in this report relate only to the items tested.  
The product complies with the requirements of the testing methods.

**Tested by:** Damjan Repar

**Approved by:** Marjan Mak



*The report shall not be reproduced except in full.*